

A new standard. A new choice.

Philips Respironics Nuance Pro gel pillows mask

Mask fitting guide



1

Before putting on the mask, place the pillow cushion tips under your nostrils. Choose the cushion size that sits comfortably under your nose without any gaps.

Important: Do not push the pillows gel base into your nostrils.



2

Pulling the headgear over your head, gently hold the pillows cushion under your nose. The headgear should lie flat against the back of your head.



3

Position the pillow cushion tips into your nostrils so that they sit comfortably under your nose without any gaps.



4

Adjust the top crown strap so that the headgear sits just above your ears and is not too close to your eyes. The headgear should fit loosely and feel comfortable.



5

Adjust the side straps evenly by pulling upward; the gel pads should sit under your cheek bones. The mask should feel secure and comfortable against your face. Do not overtighten.



6

Final adjustments for comfort and seal should be made while lying down with your device pressure applied.

Note: if excessive leaks persist, try an alternate cushion size.

Additional details on fitting may be found in the instructions that accompany your mask.

For more information, please visit www.philips.com/respironics

Broudy AE 1/14/14 MCI 4105968 PN 1114507

PHILIPS
RESPIRONICS